



<b>TEST REPORT</b>	
<b>IEC 60601-1-11</b>	
<b>MEDICAL ELECTRICAL EQUIPMENT –</b>	
<b>Part 1-11: General requirements for basic safety and essential performance –</b>	
<b>Collateral Standard: Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment</b>	
Report Reference No.....:	CHTSM19010039
Report Verification:	
Compiled by (+ signature).....:	Shelley Wang <i>shelley Wang</i>
Reviewed by (+ signature).....:	Mustang Wu <i>Mustang Wu</i>
Approved by (+ signature).....:	Tiger Jiang <i>Tiger Jiang</i>
Date of issue.....:	2019-01-24
<b>Testing Laboratory</b> .....	<b>Shenzhen Huatongwei International Inspection Co., Ltd.</b>
Address .....	Hongfa Hi-tech Industrial Park, Genyu Road, Tianliao, Gongming, Shenzhen, Guangdong, China
<b>Applicant's name</b> .....:	<b>Guangzhou Berrcom Medical Device Co., Ltd.</b>
Address .....	No.38 Huanzhen Xi Road,Dagang Town, Nansha, Guangzhou, China
<b>Test specification:</b>	
Standard .....	IEC 60601-1-11: 2015
Test procedure .....	Test Report Only
Non-standard test method.....:	N/A
<b>Test Report Form No</b> .....:	IEC60601_1_11C
Test Report Form(s) Originator .....	UL(US)
Master TRF.....:	2015-03
<b>Copyright © 2015 IEC System of Conformity Assessment Schemes for Electrotechnical Equipment and Components (IECEE System). All rights reserved.</b>	
This publication may be reproduced in whole or in part for non-commercial purposes as long as the IECEE is acknowledged as copyright owner and source of the material. IECEE takes no responsibility for and will not assume liability for damages resulting from the reader's interpretation of the reproduced material due to its placement and context.	
<b>Test item description</b> .....	<b>Non-contact Infrared Thermometer</b>
Trade Mark .....	N/A
Manufacturer .....	Guangzhou Berrcom Medical Device Co., Ltd.
Model/Type reference .....	JXB-183, JXB-178, JXB-180, JXB-186, JXB-188, JXB-190, JA001, JT001
Ratings .....	3.0Vd.c. AAA x2

**Summary of testing:****Tests performed (name of test and test clause):**

All the requirements of the standard were evaluated in this test report


**Testing location:**

Shenzhen Huatongwei International Inspection Co., Ltd.

Hongfa Hi-tech Industrial Park, Genyu Road, Tianliao, Gongming, Shenzhen, Guangdong, China

**Summary of compliance with National Differences****Copy of marking plate**

*The artwork below may be only a draft.*



<b>Test item particulars .....</b>	
Classification of installation and use .....	Hand-held
Intended use (Including type of patient, application location) .....	Refer to user manual
Mode of operation .....	Continuous / <del>non-continuous</del>
Supply Connection.....	internally powered
Accessories and detachable parts included .....	Refer to user manual
<b>Possible test case verdicts:</b>	
- test case does not apply to the test object..... :	N/A
- test object does meet the requirement .....	P (Pass)
- test object does not meet the requirement .....	F (Fail)
<b>Testing .....</b>	
Date of receipt of test item..... :	2018-12-24
Date (s) of performance of tests..... :	2019-01-15 to 2019-01-17
- Normal condition .....	N.C.
- Single fault condition .....	S.F.C.
- Means of Operator protection .....	MOOP
- Means of Patient protection .....	MOPP
<b>General remarks:</b>	
<p>"(See Enclosure #)" refers to additional information appended to the report.          "(See appended table)" refers to a table appended to the report.</p> <p>This report shall not be reproduced except in full without the written approval of the testing laboratory. List of test equipment must be kept on file and available for review. Additional test data and/or information provided in the attachments to this report.</p> <p><b>Throughout this report a <input type="checkbox"/> comma / <input checked="" type="checkbox"/> point is used as the decimal separator.</b></p> <p><b>This Test Report Form is intended for the evaluation of medical electrical equipment and medical electrical systems used in the home healthcare environment in accordance with IEC 60601-1-11.</b></p> <p><b>This Test Report Form can be used to complement the IEC 60601-1 Test Report.</b></p>	
<b>General product information:</b>	
<ol style="list-style-type: none"> <li>The Infrared Thermometer is used as medical electrical equipment;</li> <li>This thermometer is an Infrared Thermometer intended to measure forehead temperature of infants and adults without contacting human body at home or hospital, it can be used by consumers in household environment and doctor in clinic as reference.</li> <li>During the test, if the temperature is out of the accuracy range, it means the measurement is failure              Measuring range: 32.0°C ~ 43.0°C (89.2°F to 109.4°F).              Laboratory accuracy: ±0.2°C between 34°C ~ 43°C, (±0.4°F between 93.2°F ~ 109.4°F).              ±0.3°C between 32°C ~ 33.9°C, (±0.6°F between 89.2°F ~ 109.2°F).</li> <li>Max. operation temperature: 40 degree C;</li> <li>Model differences description: The circuitry of the all models is same as the model JXB-183 except the appearance and color. Unless otherwise specified, all the tests are conducted on model JXB-183 to represent other models.</li> </ol>	

IEC 60601-1-11			
Clause	Requirement + Test	Result - Remark	Verdict
<b>4</b>	<b>GENERAL REQUIREMENTS</b>		<b>P</b>
4.1	Characteristics of SUPPLY MAINS specified in 4.10.2 of Part 1 applied, except ME EQUIPMENT or ME SYSTEMS intended for HOME HEALTHCARE ENVIRONMENT complied with the following:		N/A
	– SUPPLY MAINS in the HOME HEALTHCARE ENVIRONMENT did not exceed 110 % or was not below 85 % of NOMINAL voltage between any of the conductors of the system or between any of these conductors and earth (% V) .....		—
	– For ME EQUIPMENT OR ME SYSTEMS intended to actively keep alive or resuscitate a PATIENT, SUPPLY MAINS in the HOME HEALTHCARE ENVIRONMENT did not exceed 110 % or was not below 80 % of NOMINAL voltage between any of the conductors of the system or between any of these conductors and earth (% V :	Not intend to actively keep alive or resuscitate a patient	—
	- RATED range of NOMINAL voltage did include at least 12.4 V to 15.1 V for operation from a 12 V dc supply mains	3.0Vd.c.	N/A
	- RATED range of NOMINAL voltage did include at least 24.8 V to 30.3 V for operation from a 12 V dc supply mains		N/A
	The equipment maintained BASIC SAFETY and ESSENTIAL PERFORMANCE during and following a 30 s dip to 10 V from a 12 V dc SUPPLY MAINS		N/A
	The equipment maintained BASIC SAFETY and ESSENTIAL PERFORMANCE during and following a 30 s dip to 20 V from a 24 V dc SUPPLY MAINS		N/A
<b>4.2.2</b>	<b>Environmental conditions of transport and storage between uses, indicated in instructions for use</b>		<b>P</b>
	Me equipment, except stationary equipment, after being removed from its protective packaging, and subsequently between uses, operated within its specified normal use after transport or storage in the specified environmental conditions	Indicated in the user manual: -20°C to 55°C, ≤95%RH non-condensing	P
	temperature range:-25 °C to + 5 °C		N/A
	temperature range:+5 °C to +35 °C at a non-condensing relative humidity up to 90 %		N/A
	temperature range: >35 °C to 70 °C at a water vapour pressure up to 50 hPa		N/A
	For more restricted range of environmental transport and storage conditions between uses, the environmental conditions are specified	No more restricted range is defined in user manual	N/A
	– Justified in the RISK MANAGEMENT FILE	See RISK MANAGEMENT Table 4.2.2	N/A
	– Marked on the ME EQUIPMENT		N/A
	When not practicable, the more restricted range is disclosed in the instructions for use		N/A

IEC 60601-1-11

Clause	Requirement + Test	Result - Remark	Verdict
	– Marked on the carrying case when the instructions for use indicate the ME EQUIPMENT is intended to be transported or stored in a carrying case between uses	No such required.	N/A
	Symbol 5.3.5 (ISO 7000-0534), 5.3.6 (ISO 7000-0533), or 5.3.7 (ISO 7000-0632) of ISO 15223-1:2012 used to mark temperature range		N/A
	Symbol 5.3.8 (ISO 7000-2620) of ISO 15223-1:2012 used to mark humidity range		N/A
	Symbol 5.3.9 (ISO 7000-2621) of ISO 15223-1:2012 used to mark atmospheric pressure range		N/A
	Where ME EQUIPMENT used different marking for conditions of transport and storage between uses, continuous operating conditions and transient operating conditions, markings accompanied by supplementary markings except where the respective applicability was obvious		N/A
	Environmental transport and storage test		P
	a) ME EQUIPMENT prepared for transportation or storage according to instructions for use		P
	b) ME EQUIPMENT exposed to its lowest specified environmental transport and storage conditions (temperature $\overset{0}{-4}$ °C) (°C)..... :	-24°C	P
	– For at least 16 h or, ensure ME EQUIPMENT reached THERMAL STABILITY for at least 2 h	16 h	P
	c) Then ME EQUIPMENT exposed to 34 °C ± 4 °C and 90 % - 0% + 6% relative humidity until the test chamber reached equilibrium and held for at least 2 hours. The transition from low to high temperature was made slowly enough to provide a non-condensing environment.	34°C, 90%	P
	d) ME EQUIPMENT exposed to its highest specified environmental transport and storage conditions, not requiring a water vapour pressure greater than 50 hPa (temperature $\overset{+4}{0}$ °C); (°C, ± %)..... :	59°C	P
	– For at least 16 h or, ensured ME EQUIPMENT reached THERMAL STABILITY for at least 2 h	16h	P
	e) At the end of this conditioning period, ME EQUIPMENT allowed to return and stabilize at the operating conditions of NORMAL USE		P
	f) ME EQUIPMENT evaluated to its specifications and ensured it provides BASIC SAFETY and ESSENTIAL PERFORMANCE		P
<b>4.2.3.1</b>	<b>Environmental operating conditions - Continuous operating conditions</b>		<b>P</b>
	Instructions for use indicated permissible environmental operating conditions of the ME EQUIPMENT	Indicated in the user manual: 10°C to 40°C, ≤95%RH, 70kPa to 106kPa	P

IEC 60601-1-11

Clause	Requirement + Test	Result - Remark	Verdict
--------	--------------------	-----------------	---------

	ME EQUIPMENT complied with its specifications and all requirements of the standard when operated in NORMAL USE within temperature + 5 °C to +40 °C,	Specified in User Manual: 10°C – 40°C	P
	Relative humidity range of 15 % to 90%, non-condensing, but not requiring a water vapour partial pressure greater than 50 hPa; and	humidity: ≤85%	N/A
	An atmospheric pressure range of 700 hPa to 1060 hPa	700hPa~1060hPa	P
	For more restricted range of environmental operating conditions	No more restricted range is defined in user manual	N/A
	- justified in the risk management file;		N/A
	-marked on the equipment; or were not practical in the instructions for use.....:	Refer to chapter 11 “Technical specification” in user manual	P
	– Marked on the carrying case when the instructions for use indicate the ME EQUIPMENT is intended to be operated in a carrying case	No such need.	N/A
	Symbol 5.3.5 (ISO 7000-0534), 5.3.6 (ISO 7000-0533), or 5.3.7 (ISO 7000-0632) of ISO 15223-1:2012 used to mark temperature range		N/A
	Symbol 5.3.8 (ISO 7000-2620) of ISO 15223-1:2012 used to mark humidity range		N/A
	Symbol 5.3.9 (ISO 7000-2621) of ISO 15223-1:2012 used to mark atmospheric pressure range		N/A
	Where ME EQUIPMENT used different marking for conditions of continuous operating conditions and transient operating conditions, markings accompanied by supplementary markings		N/A
	Environmental operating conditions test		P
	a) ME EQUIPMENT was set up for operation according to INTENDED USE		P
	b) ME EQUIPMENT exposed to 20 °C ± 4 °C for at least 6 h or, ensured ME EQUIPMENT reached THERMAL STABILITY for at least 2 h, (h)..... :	20°C, 6h	P
	c) ME EQUIPMENT evaluated to its specifications and ensured it continued to provide BASIC SAFETY and ESSENTIAL PERFORMANCE		P
	d) ME EQUIPMENT evaluated to its specifications and ensured it continued to provide BASIC SAFETY and ESSENTIAL PERFORMANCE while at the lowest specified atmospheric pressure.	700 hPa	P
	e) ME EQUIPMENT evaluated to its specifications and ensured it continued to provide BASIC SAFETY and ESSENTIAL PERFORMANCE while at the highest specified atmospheric pressure.	1060 hPa	P
	f) Pressure in chamber relieved		P

IEC 60601-1-11			
Clause	Requirement + Test	Result - Remark	Verdict
	g) ME EQUIPMENT cooled to its lowest specified environmental operating conditions (temperature $-4^{\circ}\text{C}$ and relative humidity less than or equal to 15 %).	6°C 15%	P
	h) ME EQUIPMENT held at lowest specified environmental operating conditions for at least 6 h or, ensured the ME EQUIPMENT reached THERMAL STABILITY for at least 2 h .....	6h	P
	i) ME EQUIPMENT met its specifications and BASIC SAFETY and ESSENTIAL PERFORMANCE		P
	j) ME EQUIPMENT warmed to its highest specified continuous environmental operating conditions (temperature $44^{\circ}\text{C}$ ).	44°C	P
	k) ME EQUIPMENT held the conditions of j) for at least 6 h or, ensured the ME EQUIPMENT reached THERMAL STABILITY for at least 2 h .....	6h	P
	l) ME EQUIPMENT met its specifications and BASIC SAFETY and ESSENTIAL PERFORMANCE		P
<b>4.2.3.2</b>	<b>Environmental shock to TRANSIT-OPERABLE EQUIPMENT</b>		N/A
	TRANSIT-OPERABLE EQUIPMENT with a stated wider range of continuous environmental operation conditions maintained BASIC SAFETY and ESSENTIAL PERFORMANCE in the presence of condensation and thermal shock from rapid changes in environmental temperature and humidity during INTENDED USE when test in accordance with 4.2.3.2 a)-j).	Not TRANSIT-OPERABLE EQUIPMENT	N/A

<b>5</b>	<b>GENERAL REQUIREMENTS FOR TESTING ME EQUIPMENT</b>		P
	In addition to the requirements of 5.9.2.1 of with IEC 60601-1 standard, accessibility determined as indicated below:		P
	ACCESSIBLE parts of ME EQUIPMENT identified by inspection and, when necessary, by testing	By inspection and testing to check.	P
	When in doubt, an ACCESSIBLE PART of ME EQUIPMENT determined by a test with the small finger probe of Fig 1, applied in a bent or straight position as follows:		N/A
	– for all positions of the ME EQUIPMENT operating in NORMAL USE		N/A
	– after opening ACCESS COVERS and removal of parts, including lamps, fuses, and fuse holders when:		N/A
	i) the ACCESS COVERS could be opened without the use of a TOOL, or		N/A
	ii) the instructions for use instructed a LAY OPERATOR to open the relevant ACCESS COVER		N/A

IEC 60601-1-11

Clause	Requirement + Test	Result - Remark	Verdict
--------	--------------------	-----------------	---------

6	CLASSIFICATION OF ME EQUIPMENT AND ME SYSTEMS		P
	ME EQUIPMENT intended for HOME HEALTHCARE ENVIRONMENT classified as follows, except for PERMANENTLY INSTALLED EQUIPMENT and as required by Part 1, Sub-clause 6.2:		P
	– CLASS II or INTERNALLY POWERED.....:	Powered by internal battery	P
	– Not provided with a FUNCTIONAL EARTH TERMINAL	No such earth terminal	P
	– When equipped with APPLIED PARTS, they are TYPE BF or CF .....	type BF applied part	P

7	ME EQUIPMENT IDENTIFICATION, MARKING AND DOCUMENTS		P
7.1	USABILITY of identification, marking, and ACCOMPANYING DOCUMENTS intended for LAY OPERATOR or LAY RESPONSIBLE ORGANIZATION evaluated by an OPERATOR whose PROFILE included minimum eight years of education	Refer to usability engineering file provided by manufacturer.	P
	ME EQUIPMENT and ME SYSTEMS intended for HOME HEALTHCARE ENVIRONMENT are simple to use and do not require referencing complex ACCOMPANYING DOCUMENTS..... :	Refer to usability engineering file provided by manufacturer.	P
7.2	In addition to requirements of 7.2.9 of the general standard, the carrying case provided some or all of the ingress protection against water or particulate matter, The ENCLOSURE is marked with the safety sign ISO 7010-W001 and “keep dry” or .....	Not use the carrying case to provide the ingress protection against water or particulate matter	N/A
	Symbol ISO 15223-1:2012, 5.3.4 (ISO 7000-0626)		N/A
	A carrying case marked with degree of protection	No such carrying case used with ME equipment	N/A
	Carrying case inspected, and tests and criteria of 7.1.2 and 7.1.3 of Part 1 applied .....		N/A
7.3	ACCOMPANYING DOCUMENTS		P
7.3.1	ACCOMPANYING DOCUMENTS indicate the LAY OPERATOR or LAY RESPONSIBLE ORGANIZATION should contact the MANUFACTURER or MANUFACTURER’S representative on the following issues:	All following information has been included in User Manual	P
	– Assistance in setting up, using, or maintaining the ME EQUIPMENT or ME SYSTEM when needed, or		P
	– To report unexpected operation or events	Refer to chapter “Troubleshooting” in user manual	P
	ACCOMPANYING DOCUMENTS include a postal address and either a phone number or web address for the LAY OPERATOR or LAY RESPONSIBLE ORGANIZATION to contact the MANUFACTURER or MANUFACTURER’S representative	See the last page in user manual	P



IEC 60601-1-11			
Clause	Requirement + Test	Result - Remark	Verdict
7.3.2	ACCOMPANYING DOCUMENTS include necessary details for healthcare professional to brief the LAY OPERATOR or LAY RESPONSIBLE ORGANIZATION on any known contraindication(s) to the use of ME EQUIPMENT or ME SYSTEM and any precautions to be taken, including the following:		P
	– Precautions to be taken in the event of changes in the performance of ME EQUIPMENT or ME SYSTEM	Refer to chapter I “Safety precautions” in user manual	P
	– Precautions to be taken regarding the exposure of the ME EQUIPMENT or ME SYSTEM to reasonably foreseeable environmental conditions	See above	P
	– Adequate information regarding medicinal substances that ME EQUIPMENT is designed to administer, including any limitations in the choice of substances to be delivered as indicated below:	No medicinal substances.	N/A
	– Information on any medicinal substances or human blood derivatives incorporated into the ME EQUIPMENT or ACCESSORIES as an integral part; and	No any medicinal substances or human blood derivatives incorporated into the ME EQUIPMENT or ACCESSORIES.	N/A
	– The degree of accuracy claimed for ME EQUIPMENT with a measuring FUNCTION		N/A
<b>7.4</b>	<b>Instructions for use</b>		P
7.4.1	Nature of the HAZARD, likely consequences that could occur if the advice is not followed, and precautions for reducing the RISK described in instructions for use corresponding to each warning and safety sign..... :	See RISK MANAGEMENT Table 7.4.1	P
	The instructions for use address the following issues, as applicable:		P
	– Strangulation due to cables and hoses, particularly due to excessive length	No such hazards	N/A
	– Inhalation or swallowing of small parts	Refer to chapter I “Safety precautions” in user manual	P
	– Potential allergic reactions to accessible materials used in the ME EQUIPMENT	No accessible allergic materials used in ME equipment	N/A
	– Contact injuries		N/A
	The instructions for use include warnings to the effect that the following actions could be unsafe as applicable:	Not other ACCESSORIES, detachable parts, or materials except Batteries	N/A
	– Use of ACCESSORIES, detachable parts, and materials not described in the instructions for use (see 7.9.2.14 of Part 1)		N/A
	– Interconnection of this equipment to other equipment not described in the instructions for use (see 16.2 c) indent 9) of Part 1)	No interconnection of this equipment to other equipment	N/A
	– Modification of the equipment	Refer to chapter XV “Explanation of symbols” in user manual.	P

IEC 60601-1-11			
Clause	Requirement + Test	Result - Remark	Verdict
	– Use of the ME EQUIPMENT outside its carrying case when some part of the protection required by this standard is provided by that carrying case (see 8.3.1 and 10.1)	No such need.	N/A
7.4.2	When BASIC SAFETY or ESSENTIAL PERFORMANCE depends on the INTERNAL ELECTRICAL POWER SOURCE, the instructions for use describes the following:	Refer to user manual	P
	– Typical operation time or number of procedures....:	Low voltage indicated	P
	– Typical service life of the INTERNAL ELECTRICAL POWER SOURCE; and.....:	No need	N/A
	– Behaviour of ME EQUIPMENT while the rechargeable INTERNAL ELECTRICAL POWER SOURCE is charging .....	Non-rechargeable battery	N/A
7.4.3	Instructions for use for ME EQUIPMENT intended for use by a LAY OPERATOR include easily understood diagrams, illustrations, or photographs of the fully assembled and ready-to-operate ME EQUIPMENT including all controls, visual INFORMATION SIGNALS, and indicators provided (see 7.1)	Refer to user manual	P
<b>7.4.4</b>	<b>Additional requirements for ME EQUIPMENT start-up PROCEDURE:</b>		P
	– Easily understood diagrams, illustrations, or photographs showing proper connection of the PATIENT to the ME EQUIPMENT, ACCESSORIES and other equipment (see 7.1)	See user manual	P
	– the time from switching “ON” until the ME EQUIPMENT is ready for NORMAL USE, when it exceeds 15 s (see 15.4.4 of Part 1) (s).....:	Less than 15s	N/A
	-the time required for ME EQUIPMENT to warm from the minimum storage temperature between uses until it is ready for intended use; and .....	Refer to user manual	P
	-the time required for ME EQUIPMENT to cool from the maximum storage temperature between uses until it is ready for intended use; and .....	Refer to user manual	P
7.4.5	Instructions for use for ME EQUIPMENT intended for use by a LAY OPERATOR include a description of generally known conditions in the HOME HEALTHCARE ENVIRONMENT that can unacceptably affect the BASIC SAFETY and ESSENTIAL PERFORMANCE of the ME EQUIPMENT	Refer to user manual	P
	The steps that can be taken by the LAY OPERATOR to identify and resolve the above conditions		P
	At least the following issues are also included as applicable		P
	- The effects of lint, dust, light (including sunlight), etc.	Refer to chapter I “Safety precautions” in user manual	P
	- A list of known devices or other sources that can potentially cause interference problems	Refer to chapter I “Safety precautions” in user manual	P
	- The effects of degraded sensors and electrodes, or loosened electrodes, that can degrade performance or cause other problems		P

IEC 60601-1-11			
Clause	Requirement + Test	Result - Remark	Verdict
	- The effects caused by pets, pests or children	Refer to chapter I "Safety precautions" in user manual	P
	The instructions for use explain the meaning of the IP classification marked on the ME EQUIPMENT, and on any carrying case provided with the ME EQUIPMENT as applicable	IP22	P
7.4.6	Instructions for use include a troubleshooting guide for use when there are indications of a ME EQUIPMENT malfunction during start-up or operation	Refer to chapter "Troubleshooting" in user manual	P
	Troubleshooting guide discloses the necessary steps in the event of an TECHNICAL ALARM CONDITION	No technical alarm.	N/A
7.4.7	Instructions for use for ME EQUIPMENT, ME SYSTEMS, parts, and ACCESSORIES for other than single use that can be contaminated by contact with PATIENT, body fluids, or expired gases, during INTENDED USE, indicate the following:		P
	– Frequency of cleaning, cleaning and disinfection, or cleaning and sterilization, as appropriate, for me equipment, me systems, parts, and accessories used on the same patient including rinsing methods, drying, handling, and storage between uses (see 8.1 and 8.2); and	Refer to chapter XI "maintenance of product" in user manual	P
	– It is necessary to clean and disinfect, clean and sterilize the ME EQUIPMENT, ME SYSTEMS, parts, and ACCESSORIES for multiple PATIENT use between uses on different PATIENTS, including rinsing methods, drying, handling, and storage until re-use (see 8.1 and 8.2), or	Special maintenance is unnecessary	N/A
	– ME EQUIPMENT, ME SYSTEMS and ACCESSORIES require professional hygienic maintenance prior to re-use and provide contact details for the source of these services (see 7.5.2)	No need professional hygienic maintenance	N/A
7.4.8	Instructions for use include:		P
	– EXPECTED SERVICE LIFE of the ME EQUIPMENT ..... :	Refer to chapter "longevity of the product" in user manual	P
	– EXPECTED SERVICE LIFE of parts and ACCESSORIES shipped with the ME EQUIPMENT ..... :	Refer to chapter VII "Features" in user manual	P
	– SHELF LIFE of parts and ACCESSORIES shipped with ME EQUIPMENT when SHELF LIFE is less than the EXPECTED SERVICE LIFE ..... :	See above	N/A
7.4.9	Instructions for use include:		P
	– A statement indicating the LAY RESPONSIBLE ORGANIZATION must contact its local authorities to determine the proper method of disposal of potentially bio hazardous parts and ACCESSORIES, as applicable	Refer to chapter XV "Explanation of symbols" in user manual.	P

IEC 60601-1-11			
Clause	Requirement + Test	Result - Remark	Verdict
7.4.10	Instructions for use includes the recommended placement of the remote parts of the DISTRIBUTED ALARM SYSTEM, when applicable, to ensure the OPERATOR can be notified at all times by an appropriate element of DISTRIBUTED ALARM SYSTEM within its specified range	No DISTRIBUTED ALARM SYSTEM.	N/A
<b>7.5</b>	<b>Technical description</b>		N/A
7.5.1	The technical description for PERMANENTLY INSTALLED CLASS I ME EQUIPMENT includes:	Not PERMANENTLY INSTALLED CLASS I ME EQUIPMENT.	N/A
	– A warning indicating the me equipment installation, including a correct protective earth connection, must only be carried out by qualified service personnel	no such warning used	N/A
	– Specifications of the PERMANENTLY INSTALLED PROTECTIVE EARTH CONDUCTOR	Not PROTECTIVE EARTH CONDUCTOR	N/A
	– A warning to verify the integrity of the external protective earthing system	Not PROTECTIVE EARTH CONDUCTOR	N/A
	– A warning to connect and verify that the PROTECTIVE EARTH TERMINAL of the PERMANENTLY INSTALLED ME EQUIPMENT is connected to the external protective earthing system	NOT PROTECTIVE EARTH TERMINAL	N/A
7.5.2	Technical description includes methods for cleaning and disinfection or cleaning and sterilization for ME EQUIPMENT and ACCESSORIES requiring professional hygienic maintenance prior to reuse (see 7.4.7):	Not require professional hygienic maintenance prior to reuse	N/A
	– Before and after any type of service PROCEDURE	See above	N/A
	– When the ME EQUIPMENT is transferred to another PATIENT	See above	N/A

<b>8</b>	<b>PROTECTION AGAINST EXCESSIVE TEMPERATURES AND OTHER HAZARDS</b>		P
8.1	A LAY OPERATOR in the HOME HEALTHCARE ENVIRONMENT can perform the cleaning or cleaning and disinfection PROCESSES when intended (see 7.4.7)	See use manual	P
	USABILITY of each such PROCESS pertaining to a LAY OPERATOR was investigated by the USABILITY ENGINEERING PROCESS .....	See usability report based on IEC 60601-1-6 provide by manufacture	P
8.2	A LAY OPERATOR in the HOME HEALTHCARE ENVIRONMENT can perform the cleaning and sterilization PROCESSES when intended (see 7.4.7)	No sterilization required	N/A
	USABILITY of each such PROCESS pertaining to a LAY OPERATOR was investigated by the USABILITY ENGINEERING PROCESS	See usability report based on IEC 60601-1-6 provide by manufacture	P
<b>8.3</b>	<b>Additional requirements for ingress of water or particulate matter into ME EQUIPMENT and ME SYSTEMS</b>		P
8.3.1	TRANSIT-OPERABLE, HANDHELD, and BODY-WORN ME EQUIPMENT maintained BASIC SAFETY and ESSENTIAL PERFORMANCE after undergoing the test of IEC 60529 for at least IP 22.....	Hand-held equipment, IP22	P

IEC 60601-1-11			
Clause	Requirement + Test	Result - Remark	Verdict
	All other ME EQUIPMENT maintained BASIC SAFETY and ESSENTIAL PERFORMANCE after undergoing the test of IEC 60529 for at least IP21..... :	IP22	N/A
	For PORTABLE ME EQUIPMENT intended to be used only while in a carrying case, IP21 met with the ME EQUIPMENT in its the carrying case	See above	N/A
	Maintenance of BASIC SAFETY and ESSENTIAL PERFORMANCE VERIFIED	Verified	P
8.3.2	ENCLOSURES of the non-ME EQUIPMENT parts of the ME SYSTEMS provide the degree of protection against harmful ingress of water or particulate matter equivalent to equipment complying with their respective IEC or ISO safety standards	Not ME SYSTEMS	N/A
	Tests of IEC 60529:1989 conducted with the equipment placed in the least favourable position of NORMAL USE and the ENCLOSURES inspected		N/A
<b>8.4</b>	<b>Additional requirements for interruption of the power supply/SUPPLY MAINS to ME EQUIPMENT and ME SYSTEM</b>		N/A
	ME EQUIPMENT or ME SYSTEM with ESSENTIAL PERFORMANCE intended to actively keep alive or resuscitate a PATIENT maintained its ESSENTIAL PERFORMANCE for a sufficient time or for a sufficient number of PROCEDURES when loss or failure of SUPPLY MAINS or near depletion INTERNAL ELECTRICAL POWER SOURCE occurred	Not such ME EQUIPMENT	N/A
	The time or number of PROCEDURES remaining allowed alternative life-supporting methods to be employed		N/A
	Optionally, an INTERNAL ELECTRICAL POWER SOURCE was used to maintain ESSENTIAL PERFORMANCE ..... :		N/A
	Optionally, independent means were used to provide ESSENTIAL PERFORMANCE..... :		N/A
	Instructions for use disclose the time or number of procedures available following a loss or failure of the SUPPLY MAINS or near depletion of the INTERNAL ELECTRICAL POWER SOURCE		N/A
	Instructions for use describes the alternative life-supporting methods to be employed		N/A
	The technical description describes methods that can be employed for longer periods		N/A
	ME EQUIPMENT or ME SYSTEM with ESSENTIAL PERFORMANCE intended to actively keep alive or resuscitate a PATIENT with no INTERNAL ELECTRICAL POWER SOURCE is equipped with an ALARM SYSTEM that includes at least a MEDIUM PRIORITY ALARM CONDITION indicating power failure ..... :		N/A
	ME EQUIPMENT or ME SYSTEM with ESSENTIAL PERFORMANCE intended to actively keep alive or resuscitate a PATIENT with an INTERNAL ELECTRICAL POWER SOURCE is equipped with an automatic switchover to INTERNAL ELECTRICAL POWER SOURCE		N/A

IEC 60601-1-11			
Clause	Requirement + Test	Result - Remark	Verdict
	ME EQUIPMENT or ME SYSTEM with ESSENTIAL PERFORMANCE intended to actively keep alive or resuscitate a PATIENT with an INTERNAL ELECTRICAL POWER SOURCE is equipped with an ALARM SYSTEM that includes at least a MEDIUM PRIORITY TECHNICAL ALARM CONDITION indicating the INTERNAL ELECTRICAL POWER SOURCE is nearing insufficient remaining power for operation		N/A
	TECHNICAL ALARM CONDITION provides sufficient time or sufficient number of procedures for a LAY OPERATOR to act		N/A
	A TECHNICAL ALARM CONDITION of at least LOW PRIORITY remained active until the INTERNAL ELECTRICAL POWER SOURCE returned to a level above the ALARM LIMIT or until depleted		N/A
	It was not possible to inactivate the visual ALARM SIGNAL of this TECHNICAL ALARM CONDITION		N/A
	Functional tests conducted, and the RISK MANAGEMENT FILE inspected..... :		N/A
<b>8.5</b>	<b>Additional requirements for an INTERNAL ELECTRICAL POWER SOURCE</b>		<b>P</b>
8.5.1	ME EQUIPMENT provided with a means for the OPERATOR to determine state of the INTERNAL ELECTRICAL POWER SOURCE when the is essential for BASIC SAFETY or ESSENTIAL PERFORMANCE or to control risks associated with loss of ESSENTIAL PERFORMANCE		P
	State of INTERNAL ELECTRICAL POWER SOURCE indicated by:		P
	- number of PROCEDURES remaining;		N/A
	-remaining operating time;		N/A
	-percentage of the remaining operating time or energy; or		P
	-"fuel" gauge		N/A
	Instructions described method to determine state of INTERNAL ELECTRICAL POWER SOURCE	Refer to chapter "CHARGINH THE BATTRYS" in user manual	P
8.5.2	Means, other than labelling, provided to prevent RISK of swallowing coin/button cells	No button cell used in ME equipment	N/A
	Replacement of button cell require use of TOOL		N/A

<b>9</b>	<b>ACCURACY OF CONTROLS AND INSTRUMENTS AND PROTECTION AGAINST HAZARDOUS OUTPUTS</b>		<b>P</b>
	The RISKS associated with USABILITY in the HOME HEALTHCARE ENVIRONMENT for OPERATOR PROFILES including a LAY OPERATOR when performing the USABILITY ENGINEERING PROCESS include at least the following considerations:		P
	- changes of controls		N/A
	- unexpected movement	Hand-held equipment	N/A
	- potential for misconnection		N/A

IEC 60601-1-11			
Clause	Requirement + Test	Result - Remark	Verdict
	– potential for improper operation, or unsafe use		P
	– potential for confusion as to current operational mode		N/A
	– change in the transfer of energy or substance	No change in the transfer or substance	N/A
	- exposure to environmental conditions specified in this standard		N/A
	– exposure to biological materials, and		N/A
	– small parts being inhaled or swallowed		P
	Particular emphasis placed on the limited training of a LAY OPERATOR with respect to the ability to intervene and maintain BASIC SAFETY and ESSENTIAL PERFORMANCE.	See user manual	P
	The MANUFACTURER'S USABILITY ENGINEERING PROCESS included the least capable intended LAY OPERATOR or LAY RESPONSIBLE ORGANIZATION	Refer to usability engineering file provided by manufacturer.	P
	USABILITY ENGINEERING FILE inspected for compliance .....	Refer to usability engineering file provided by manufacturer.	P

<b>10</b>	<b>CONSTRUCTION OF ME EQUIPMENT</b>		P
10.1	Additional requirements for mechanical strength		P
10.1.1	Additions to Table 28 Mechanical strength test of the base standard, conducted as indicated in Table 1, Mechanical strength test applicability, non-TRANSIT-OPERABLE, and Table 2, Mechanical strength test applicability, TRANSIT-OPERABLE	non-TRANSIT-OPERABLE	P
10.1.2	ME EQUIPMENT, its parts, and mounting ACCESSORIES, intended for non-TRANSIT-OPERABLE use displayed adequate mechanical strength when subjected to mechanical stress caused by NORMAL USE, including pushing, impact, dropping and rough handling (not applicable to FIXED and STATIONARY ME EQUIPMENT)	Hand-held equipment, and intended for non-TRANSIT-OPERABLE use.	P
	ME EQUIPMENT maintained BASIC SAFETY and ESSENTIAL PERFORMANCE after mechanical tests		P
	OPERATOR-re-settable protective devices that can be reset without the use of a TOOL were, optionally, reset prior to the evaluation of BASIC SAFETY and ESSENTIAL PERFORMANCE		N/A
	a) Shock tests conducted in accordance with IEC 60068-2-27:2008 .....	See Appended Table 10.1.2a	P
	b) Broad-band random vibration tests conducted in accordance with IEC 60068-2-64:2008, using the following conditions .....	See Appended Table 10.1.2b	P

IEC 60601-1-11			
Clause	Requirement + Test	Result - Remark	Verdict
10.1.3	ME EQUIPMENT, parts, and mounting ACCESSORIES for TRANSIT-OPERABLE use displayed adequate mechanical strength when subjected to pushing, impact, dropping, rough handling, and rigorous conditions of PATIENT movement in NORMAL USE as well as transportation by trolleys, carts, road vehicles, trains, ships, and aircraft	Non-TRANSIT-OPERABLE, not used at the conditions of PATIENT movement in NORMAL USE.	N/A
	ME EQUIPMENT maintained BASIC SAFETY and ESSENTIAL PERFORMANCE after the following tests:		N/A
	a) Shock tests conducted on other than HAND-HELD ME EQUIPMENT, parts, and mounting ACCESSORIES in accordance with IEC 60068-2-27:2008	HAND-HELD ME EQUIPMENT	N/A
	1) Test type: Type 1 .....		N/A
	2) Test type: Type 2 .....		N/A
	b) Shock tests conducted on HAND-HELD ME EQUIPMENT, parts, and mounting ACCESSORIES in accordance with IEC 60068-2-27:2008	HAND-HELD ME EQUIPMENT	N/A
	1) Test type: Type 1 .....		N/A
	2) Test type: Type 2 .....		N/A
	c) Broad-band random vibration test conducted on ME EQUIPMENT, parts, and mounting ACCESSORIES in accordance with IEC 60068-2-64:2008 .....	non-TRANSIT-OPERABLE	N/A
	d) Free fall tests conducted on PORTABLE and MOBILE ME EQUIPMENT, parts, and mounting ACCESSORIES per IEC 60068-2-31:2008, using PROCEDURE 1 .....	non-TRANSIT-OPERABLE	N/A
	BASIC SAFETY and ESSENTIAL PERFORMANCE were maintained		N/A
10.2	Controls of ME EQUIPMENT intended for use by a LAY OPERATORY that can affect BASIC SAFETY or ESSENTIAL PERFORMANCE protected from accidental or unauthorized changes or adjustments		N/A
	OPERATOR-adjustable controls used for calibration include a means to prevent unintentional changes from the intended position	No OPERATOR-adjustable control.	N/A
<b>11</b>	<b>PROTECTION AGAINST STRANGULATION OR ASPHYXIATION</b>		N/A
	Means provided to control the RISK of strangulation and asphyxiation of the PATIENT and others to an acceptable level	No such risk	N/A
	EQUIPMENT and RISK MANAGEMENT FILE inspected .....		N/A
<b>12</b>	<b>ADDITIONAL REQUIREMENTS FOR ELECTROMAGNETIC EMISSIONS OF ME EQUIPMENT AND ME SYSTEMS</b>		P
	ME EQUIPMENT and ME SYSTEMS intended for HOME HEALTHCARE ENVIRONMENT are Class B according to CISPR 11:2009 .....		P



IEC 60601-1-11			
Clause	Requirement + Test	Result - Remark	Verdict
<b>13</b>	<b>ADDITIONAL REQUIREMENTS FOR ALARM SYSTEMS OF ME EQUIPMENT AND ME SYSTEMS</b>		N/A
13.1	Each HIGH PRIORITY and MEDIUM PRIORITY ALARM CONDITION causes generation of auditory ALARM SIGNALS per IEC 60601-1-8:2006 and IEC 60601-1-8:2006/AMD1:2012, except when equipment is connected to a DISTRIBUTED ALARM SYSTEM intended for confirmed deliver of ALARM CONDITIONS including the generation of auditory ALARM SIGNALS per IEC 60601-1-8:2006 and IEC 60601-1-8:2006/AMD1:2012 .....	No ALARM SYSTEM.	N/A
13.2	For ME EQUIPMENT and ME SYSTEMS intended to actively keep alive or resuscitate a PATIENT, reducing the auditory ALARM SIGNAL volume T below audible levels resulted in the following was not possible, except when the ALARM SYSTEM was connected to a DISTRIBUTED ALARM SYSTEM that included generation of auditory ALARM SIGNALS per IEC 60601-1-8:2006 and IEC 60601-1-8:2006/AMD1:2012		N/A

<b>4.2.2</b>	<b>RM RESULTS TABLE: Permissible environmental conditions of transport and storage, between uses, indicated in instructions for use</b>		<b>N/A</b>
<b>Clause of ISO 14971</b>	<b>Document Ref. in RMF (Document No. &amp; paragraph)</b>	<b>Result - Remarks</b>	<b>Verdict</b>
4.2			
4.3			
4.4			

<b>4.2.3.1</b>	<b>RM RESULTS TABLE: Environmental operating conditions - Continuous operating conditions</b>		<b>N/A</b>
<b>Clause of ISO 14971</b>	<b>Document Ref. in RMF (Document No. &amp; paragraph)</b>	<b>Result - Remarks</b>	<b>Verdict</b>
4.2			
4.3			
4.4			

<b>7.4.1</b>	<b>RM RESULTS TABLE: Additional requirements for warning and safety notices</b>		<b>P</b>
<b>Clause of ISO 14971</b>	<b>Document Ref. in RMF (Document No. &amp; paragraph)</b>	<b>Result - Remarks</b>	<b>Verdict</b>
4.2	RAR-File No.: BER-RD-005/A0 Appended Table 1	Intended use, Identification of characteristics related to safety	P
4.3	RAR-File No.: BER-RD-005/A0 Appended Table 2	Identification of hazards	P
4.4	RAR-File No.: BER-RD-005/A0 Appended Table 3	Estimation of the risks	P
5	RAR-File No.: BER-RD-005/A0 Appended Table 3	Risk evaluation	P
6.2	RAR-File No.: BER-RD-005/A0 Appended Table 3	Risk control	P

<b>7.4.5</b>	<b>RM RESULTS TABLE: : Additional requirements for operating instructions</b>		<b>P</b>
<b>Clause of ISO 14971</b>	<b>Document Ref. in RMF (Document No. &amp; paragraph)</b>	<b>Result - Remarks</b>	<b>Verdict</b>
4.3	RAR-File No.: BER-RD-005/A0 Appended Table 2	Identification of hazards	P
4.4	RAR-File No.: BER-RD-005/A0 Appended Table 3	Estimation of the risks	P
5	RAR-File No.: BER-RD-005/A0 Appended Table 3	Risk evaluation	P
6.2	RAR-File No.: BER-RD-005/A0 Appended Table 3	Risk control	P

<b>8.4</b>	<b>RM RESULTS TABLE: Additional requirements for interruption of power supply / supply mains to ME Equipment and ME Systems</b>		<b>N/A</b>
------------	---	--	------------

IEC 60601-1-11
----------------

Clause of ISO 14971	Document Ref. in RMF (Document No. & paragraph)	Result - Remarks	Verdict
4.2			
4.3			
5			
6.2			
6.3			
6.4			
6.5			
6.6			
6.7			

<b>10.1.2a</b>	<b>TABLE: Shock test (IEC 60068-2-27:2008), using the following conditions*:</b>	<b>P</b>
	Peak acceleration ..... :	150 m/s <sup>2</sup> (15 g)
	Duration..... :	11 ms
	Pulse shape ..... :	half-sine
	Number of shocks..... :	3 shocks per direction per axis (18 total)

Direction Shock Applied	Axis Shock Applied	BASIC SAFETY and ESSENTIAL PERFORMANCE maintained? Yes/No	Remarks
+X	X	Yes	No damage.
+X		Yes	No damage.
+X		Yes	No damage.
-X		Yes	No damage.
-X		Yes	No damage.
-X		Yes	No damage.
+Y	Y	Yes	No damage.
+Y		Yes	No damage.
+Y		Yes	No damage.
-Y		Yes	No damage.
-Y		Yes	No damage.
-Y		Yes	No damage.
+Z	Z	Yes	No damage.
+Z		Yes	No damage.
+Z		Yes	No damage.
-Z		Yes	No damage.
-Z		Yes	No damage.
-Z		Yes	No damage.

Supplementary information:  
 \*(NOTE 1 This represents Class 7M1 as described in IEC TR 60721-4-7:2001 [6])



IEC 60601-1-11
----------------


Supplementary information:  
 \* (NOTE 3 This represents Class 7M2 as described in IEC/TR 60721-4-7:2001 [6])

<b>10.1.3a2</b>	<b>TABLE: Shock test (IEC 60068-2-27:2008) on other than HAND-HELD ME EQUIPMENT, parts, and mounting ACCESSORIES under the following conditions (Test Type 2):</b>	<b>N/A</b>
-----------------	--	------------

	Peak acceleration .....	300 m/s <sup>2</sup> (15 g)
	Duration .....	6 ms
	Pulse shape .....	half-sine
	Number of shocks .....	3 shocks per direction per axis (18 total)

Direction Shock Applied	Axis Shock Applied	BASIC SAFETY and ESSENTIAL PERFORMANCE maintained? Yes/No	Remarks

Supplementary information:

<b>10.1.3b1</b>	<b>TABLE: Shock test (IEC 60068-2-27:2008) on HAND-HELD ME EQUIPMENT parts, and mounting ACCESSORIES using the following conditions (Test Type 1):</b>	<b>N/A</b>
-----------------	--	------------

	Peak acceleration .....	300 m/s <sup>2</sup> (30 g)
	Duration .....	11 ms
	Pulse shape .....	half-sine
	Number of shocks .....	3 shocks per direction per axis (18 total)

Direction Shock Applied	Axis Shock Applied	BASIC SAFETY and ESSENTIAL PERFORMANCE maintained? Yes/No	Remarks

IEC 60601-1-11


Supplementary information:  
 \*(NOTE 4 This represents Class 7M3 as described in IEC/TR 60721-4-7:2001. (Test Type 1))

<b>10.1.3b2</b>	<b>TABLE: Shock test (IEC 60068-2-27:2008) on HAND-HELD ME EQUIPMENT parts, and mounting ACCESSORIES using the following conditions (Test Type 2):</b>	<b>N/A</b>
	Peak acceleration .....	1000 m/s <sup>2</sup> (100 g)
	Duration .....	6 ms
	Pulse shape .....	half-sine
	Number of shocks .....	3 shocks per direction per axis (18 total)

Direction Shock Applied	Axis Shock Applied	BASIC SAFETY and ESSENTIAL PERFORMANCE maintained? Yes/No	Remarks

Supplementary information:

IEC 60601-1-11

<b>10.1.3c</b>	<b>TABLE: Broad-band random vibration test (IEC 60068-2-64:2008) on ME EQUIPMENT, parts, and mounting ACCESSORIES using the following conditions*:</b>		<b>N/A</b>
1	Acceleration amplitude.....:	10 Hz to 100 Hz: 1,0 (m/s <sup>2</sup> ) <sup>2</sup> /Hz	
2	Acceleration amplitude.....:	100 Hz to 200 Hz: - 3 db per octave	
3	Acceleration amplitude.....:	200 Hz to 2 000 Hz: 0,5 (m/s <sup>2</sup> ) <sup>2</sup> /Hz	
	Duration .....	30 min per perpendicular axis (3 total)	
Perpendicular axis subjected to broad-band random vibration test	Acceleration amplitude	BASIC SAFETY and ESSENTIAL PERFORMANCE maintained? Yes/No	Remarks
1	1		
2	1		
3	1		
1	2		
2	2		
3	2		
1	3		
2	3		
3	3		
Supplementary information:			
*(NOTE 5 This represents Class 7M1 and 7M2 as described in IEC/TR 60721-4-7:2001)			

<b>10.1.3d</b>	<b>TABLE: Free fall test (IEC 60068-2-31:2008), using PROCEDURE 1, on PORTABLE and MOBILE ME EQUIPMENT, parts, and mounting ACCESSORIES (with carrying case if intended), under the following conditions*:</b>		<b>N/A</b>	
1	Fall height for mass ≤ 1 kg.....:	0,25 m		
2	Fall height for mass > 1 kg and ≤ 10 Kg.....:	0,1 m		
3	Fall height for mass > 10 kg and ≤ 50 Kg.....:	0,05 m		
4	Fall height for mass > 50 kg.....:	0,01 m		
Specified altitude (m)	Mass (Kg)	Fall No.	BASIC SAFETY and ESSENTIAL PERFORMANCE maintained? Yes/No	Remarks
0,25	≤ 1	1		
0,25	≤ 1	2		
0,1	> 1 & ≤ 10	1		
0,1	> 1 & ≤ 10	2		
0,05	> 10 & ≤ 50	1		
0,05	> 10 & ≤ 50	2		
0,01	> 50	1		
0,01	> 50	2		
Supplementary information:				
*(NOTE 6 This represents Class 7M2 as described in IEC/TR 60721-4-7:2001)				

11.0	RM RESULTS TABLE: PROTECTION AGAINST STRANGULATION AND ASPHYXIATION		N/A
Clause of ISO 14971	Document Ref. in RMF (Document No. & paragraph)	Result - Remarks	Verdict
4.3			
4.4			
5			
6.2			
6.3			
6.4			
6.5			
6.6			
Supplementary information:			

--- End of Test Report ---